

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/04/2008 08:26

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2950
Check sum	74790	Download date/time:	01/04/2008	08:27		Record ID	3B97
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP	COBRA		

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/04/2008 08:31

First calibration check: .130

Interferent detected: No

Second calibration check: .131

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS052807

Soln temp C: 34.03

Simulator Ser No. DR3707

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards			
Prom	Real Time clock	Motor	Printer			
Ram	Auto calibration	Range	Digital voltmeter			
Check sum	48949	Download date/time:	01/04/2008	08:31	Record ID	3BA6
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP	COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/10/2008 10:13

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2912
Check sum	51678	Download date/time:	01/10/2008	10:13		Record ID	3BK6
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP	COBRA		

Instrument serial No: 68-012336

Accuracy check date/time: 01/10/2008 10:22

First calibration check: .133

Interferent detected: No

Second calibration check: .130

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS052807

Soln temp C: 34.00

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	22837	Download date/time:	01/10/2008	10:22		Record ID	3BL7
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP	COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/16/2008 13:57

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2937
Check sum	52435	Download date/time:	01/16/2008	13:57		Record ID	3BS3
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/16/2008 14:00

First calibration check: .129

Interferent detected: No

Second calibration check: .129

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS052807

Soln temp C: 34.01

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	63561	Download date/time:	01/16/2008	14:00		Record ID	3BT8
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/25/2008 10:09

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2841
Check sum	57752	Download date/time:	01/25/2008	10:09		Record ID	3C16
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP	COBRA		

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/25/2008 10:12

First calibration check: .132

Interferent detected: No

Second calibration check: .135

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS052807

Soln temp C: 34.01

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 004

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	45701	Download date/time:	01/25/2008	10:12		Record ID	3C27
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP	COBRA		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/31/2008 12:23

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2913
Check sum	65148	Download date/time:	01/31/2008	12:27		Record ID	3CL6
	Record imported (Y/N)	Y	Source Diagnostic				:LAPTOP COBRA

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/31/2008 12:25

First calibration check: .128

Interferent detected: No

Second calibration check: .129

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS052807

Soln temp C: 0000

Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 004

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	28695	Download date/time:	01/31/2008	12:27		Record ID	3CJ4
	Record imported (Y/N)	Y	Source Calibration				:LAPTOP COBRA

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/31/2008 12:27

First calibration check: .130

Interferent detected: No

Second calibration check: .131

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS052807

Soln temp C: 34.03  
Simulator Ser No. DR3707

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57098	Download date/time: 01/31/2008	12:27	Record ID 3CK7
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.