

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/05/2008 12:19

First calibration check: .120

Interferent detected: No

Second calibration check: .121

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.00
Simulator Ser No. DR3707

PLACED INTO SERVICE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 10000	Download date/time: 03/14/2008	19:40	Record ID 3EH2
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012335 Accuracy check date/time: 03/05/2008 12:21

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3723
Check sum 22760	Download date/time: 03/14/2008	19:40	Record ID 3E19				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 03/05/2008 13:23

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3744
Check sum	50731	Download date/time:	03/14/2008 19:40	Record ID	3EJ6		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012335

Accuracy check date/time: 03/07/2008 08:08

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3554
Check sum	59434	Download date/time:	03/14/2008 19:40	Record ID	3EK3		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/08/2008 03:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3561
Check sum	69410	Download date/time:	03/14/2008	19:40		Record ID	3EL4
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012335 Accuracy check date/time: 03/08/2008 03:04

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	38496	Download date/time:	03/14/2008	19:40		Record ID	3EP3
Record imported (Y/N)		Y	Source Exception		:LAPTOP COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/09/2008 04:03

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16031	Download date/time: 03/14/2008	19:40	Record ID 3EQ6
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012335 Accuracy check date/time: 03/09/2008 04:05

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 60428	Download date/time: 03/14/2008	19:40	Record ID 3ER7
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/09/2008 04:07

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 21090	Download date/time: 03/14/2008	19:40	Record ID 3ES0
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012335 Accuracy check date/time: 03/10/2008 01:30

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No. Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3527
Check sum 24482	Download date/time: 03/14/2008	19:40	Record ID	3EM1			
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/10/2008 08:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

CHANGED TIME FROM 0758 TO 0858

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16389	Download date/time: 03/14/2008	19:40	Record ID 3ET5
Record imported (Y/N) Y	Source Memo	:LAPTOP COBRA	

Instrument serial No: 68-012335 Accuracy check date/time: 03/10/2008 08:59

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No. Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3481
Check sum 62839	Download date/time: 03/14/2008	19:40	Record ID	3EN8			
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 03/14/2008 19:39

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3505
Check sum	18411	Download date/time:	03/14/2008	19:40		Record ID	3E05
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP	COBRA		

Instrument serial No: 68-012335

Accuracy check date/time: 03/14/2008 19:52

First calibration check: .116

Interferent detected: No

Second calibration check: .115

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.02

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards			
Prom	Real Time clock	Motor	Printer			
Ram	Auto calibration	Range	Digital voltmeter			
Check sum	30795	Download date/time:	03/14/2008	19:53	Record ID	3F42
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP	COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 03/14/2008 20:43

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 45780	Download date/time: 03/20/2008	12:16	Record ID 3FI8
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012335

Accuracy check date/time: 03/18/2008 11:48

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3522
Check sum 39356	Download date/time: 03/20/2008	12:16	Record ID 3FG6				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/20/2008 12:09

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3511
Check sum	26051	Download date/time:	03/20/2008	12:16		Record ID	3FH1
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012335 Accuracy check date/time: 03/20/2008 12:23

First calibration check: .118

Interferent detected: No

Second calibration check: .118

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	49131	Download date/time:	03/20/2008	12:23		Record ID	3FK8
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 03/26/2008 14:52

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 20886	Download date/time: 03/27/2008	08:33	Record ID 3FQ1
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012335

Accuracy check date/time: 03/27/2008 01:11

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 61400	Download date/time: 03/27/2008	08:33	Record ID 3FR8
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 03/27/2008 08:31

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3507
Check sum	40478	Download date/time:	03/27/2008	08:33		Record ID	3FP8
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012335 Accuracy check date/time: 03/27/2008 08:36

First calibration check: .115

Interferent detected: No

Second calibration check: .115

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.04

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 034

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	65184	Download date/time:	03/27/2008	08:37		Record ID	3FZ4
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.