

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334

Accuracy check date/time: 04/04/2007 10:54

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3156
Check sum	40992	Download date/time:	04/04/2007	11:00		Record ID	2TV7
	Record imported (Y/N)	Y	Source	Diagnostic		:LAPTOP	COBRA

Instrument serial No: 68-012334

Accuracy check date/time: 04/04/2007 10:56

First calibration check: .134

Interferent detected: No

Second calibration check: .138

Interferent detected: No

Concentration (%W/V) .140

Solution lot#: IRS091906

Soln temp C: 33.99

Simulator Ser No. DR3747

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	32503	Download date/time:	04/04/2007	11:00		Record ID	2TT7
	Record imported (Y/N)	Y	Source	Calibration		:LAPTOP	COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 04/04/2007 10:59

First calibration check: .139

Interferent detected: No

Second calibration check: .140

Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS091906

Soln temp C: 34.00
Simulator Ser No. DR3747

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 52862	Download date/time: 04/04/2007	11:00	Record ID 2TU4
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012334 Accuracy check date/time: 04/04/2007 11:01

First calibration check: .140

Interferent detected: No

Second calibration check: .141

Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS091906

Soln temp C: 33.99
Simulator Ser No. DR3747

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 37212	Download date/time: 04/04/2007	11:02	Record ID 2TW4
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334

Accuracy check date/time: 04/12/2007 07:19

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3210
Check sum	34033	Download date/time:	04/18/2007	11:51		Record ID	2UO2
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012334

Accuracy check date/time: 04/12/2007 07:20

First calibration check: .134

Interferent detected: No

Second calibration check: .138

Interferent detected: No

Concentration (%W/V) .140

Solution lot#: IRS091906

Soln temp C: 34.03

Simulator Ser No. DR3747

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	29996	Download date/time:	04/18/2007	11:51		Record ID	2UL9
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334

Accuracy check date/time: 04/15/2007 05:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3393
Check sum	19684	Download date/time:	04/18/2007 11:51	Record ID	2UP8		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012334

Accuracy check date/time: 04/18/2007 11:46

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3155
Check sum	52678	Download date/time:	04/18/2007 11:51	Record ID	2UQ1		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 04/18/2007 11:48

First calibration check: .130

Interferent detected: No

Second calibration check: .137

Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS091906

Soln temp C: 33.05
Simulator Ser No. DR3747

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 60618	Download date/time: 04/18/2007	11:51	Record ID 2UM6
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012334 Accuracy check date/time: 04/18/2007 11:50

First calibration check: .140

Interferent detected: No

Second calibration check: .141

Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS091906

Soln temp C: 34.01
Simulator Ser No. DR3747

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 38717	Download date/time: 04/18/2007	11:51	Record ID 2UN5
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334

Accuracy check date/time: 04/27/2007 00:35

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3175
Check sum	38279	Download date/time:	04/27/2007	00:39		Record ID	2VD8
	Record imported (Y/N)	Y	Source Diagnostic			:LAPTOP	COBRA

Instrument serial No: 68-012334

Accuracy check date/time: 04/27/2007 00:37

First calibration check: .133

Interferent detected: No

Second calibration check: .137

Interferent detected: No

Concentration (%W/V) .140

Solution lot#: IRS091906

Soln temp C: 34.01

Simulator Ser No. DR3747

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	57547	Download date/time:	04/27/2007	00:39		Record ID	2VC1
	Record imported (Y/N)	Y	Source Calibration			:LAPTOP	COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.