

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336

Accuracy check date/time: 06/02/2008 09:40

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2820
Check sum	13500	Download date/time:	06/02/2008	09:43		Record ID	3J81
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012336

Accuracy check date/time: 06/02/2008 09:42

First calibration check: .110

Interferent detected: No

Second calibration check: .112

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 003

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards			
Prom	Real Time clock	Motor	Printer			
Ram	Auto calibration	Range	Digital voltmeter			
Check sum	18248	Download date/time:	06/02/2008	09:43	Record ID	3J74
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336 Accuracy check date/time: 06/02/2008 09:46

First calibration check: .114

Interferent detected: No

Second calibration check: .115

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.05
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16716	Download date/time: 06/02/2008	09:47	Record ID 3J92
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012336 Accuracy check date/time: 06/06/2008 12:43

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2872
Check sum 21205	Download date/time: 06/06/2008	12:47	Record ID 3K33				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336 Accuracy check date/time: 06/06/2008 12:47

First calibration check: .114

Interferent detected: No

Second calibration check: .115

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 0000
Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 98840	Download date/time: 06/06/2008	12:47	Record ID 3K24
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012336 Accuracy check date/time: 06/08/2008 21:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 67920	Download date/time: 06/13/2008	12:45	Record ID 3K59
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336 Accuracy check date/time: 06/08/2008 21:51

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 34450	Download date/time: 06/13/2008	12:45	Record ID 3K64
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012336 Accuracy check date/time: 06/13/2008 12:44

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2808
Check sum 30564	Download date/time: 06/13/2008	12:45	Record ID 3K40				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336 Accuracy check date/time: 06/13/2008 12:48

First calibration check: .116

Interferent detected: No

Second calibration check: .116

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 005

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30675	Download date/time: 06/13/2008	12:48	Record ID 3K73
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012336 Accuracy check date/time: 06/20/2008 09:35

First calibration check: .112

Interferent detected: No

Second calibration check: .113

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.03
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator 7851

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 35925	Download date/time: 07/03/2008	14:53	Record ID 3LC1
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336 Accuracy check date/time: 06/20/2008 09:40

First calibration check: .114

Interferent detected: No

Second calibration check: .114

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator 7851

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 19569	Download date/time: 07/03/2008	14:53	Record ID 3LD0
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012336 Accuracy check date/time: 06/27/2008 15:55

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2700
Check sum 16256	Download date/time: 07/03/2008	14:53	Record ID 3LF0				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012336

Accuracy check date/time: 06/27/2008 15:58

First calibration check: .117

Interferent detected: No

Second calibration check: .119

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 15380	Download date/time: 07/03/2008	14:53	Record ID 3LE9
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.