

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 12/05/2008 13:03

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3394
Check sum	14823	Download date/time:	12/05/2008	13:04		Record ID	3Y61
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

---

Instrument serial No: 68-012334 Accuracy check date/time: 12/05/2008 13:05

First calibration check: .116

Interferent detected: No

Second calibration check: .118

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards		
Prom	Real Time clock	Motor	Printer		
Ram	Auto calibration	Range	Digital voltmeter		
Check sum	57171	Download date/time:	12/05/2008 13:04	Record ID	3Y58
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 12/12/2008 10:43

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3350
Check sum	30964	Download date/time:	12/12/2008	10:44		Record ID	3Y78
	Record imported (Y/N)	Y	Source	Diagnostic		:LAPTOP	COBRA

---

Instrument serial No: 68-012334 Accuracy check date/time: 12/12/2008 10:48

First calibration check: .110

Interferent detected: No

Second calibration check: .112

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.04

Simulator Ser No. DR3704

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	59374	Download date/time:	12/12/2008	10:50		Record ID	3Y81
	Record imported (Y/N)	Y	Source	Calibration		:LAPTOP	COBRA

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 12/12/2008 10:50

First calibration check: .114

Interferent detected: No

Second calibration check: .116

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.01  
Simulator Ser No. DR3704

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 23093	Download date/time: 12/12/2008	10:50	Record ID 3Y98
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 12/19/2008 09:33

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3293
Check sum 64288	Download date/time: 12/19/2008	09:34	Record ID 3YJ4
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 12/19/2008 09:46

First calibration check: .115

Interferent detected: No

Second calibration check: .114

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.04  
Simulator Ser No. DR3747

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51542	Download date/time: 12/19/2008	09:46	Record ID 3YK1
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 12/26/2008 11:12

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3284
Check sum 51940	Download date/time: 12/26/2008	11:12	Record ID 3Z73
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334

Accuracy check date/time: 12/26/2008 11:14

First calibration check: .114

Interferent detected: No

Second calibration check: .116

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.02

Simulator Ser No. DR3747

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 14863	Download date/time: 12/26/2008	11:14	Record ID 3Z80
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.