

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/02/2009 08:42

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3689
Check sum	58069	Download date/time:	01/02/2009 08:46	Record ID	3ZA4		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012321 Accuracy check date/time: 01/02/2009 08:45

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3691
Check sum	50095	Download date/time:	01/02/2009 08:46	Record ID	3ZB3		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/02/2009 08:47

First calibration check: .121

Interferent detected: No

Second calibration check: .121

Interferent detected: No

Concentration (%W/V) .124 Solution lot#: IRS020708

Soln temp C: 34.02
Simulator Ser No. DR2492

Number of people tested since last periodic determination of accuracy: 041

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 58680	Download date/time: 01/02/2009	08:48	Record ID 3ZC0
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/03/2009 01:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16818	Download date/time: 01/09/2009	09:38	Record ID 3ZL7
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/09/2009 09:33

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3704
Check sum	61111	Download date/time:	01/09/2009	09:38		Record ID	3ZK0
		Record imported (Y/N)	Y	Source Diagnostic			:LAPTOP COBRA

Instrument serial No: 68-012321

Accuracy check date/time: 01/09/2009 09:38

First calibration check: .136

Interferent detected: No

Second calibration check: .138

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.05

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 023

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	11424	Download date/time:	01/09/2009	09:38		Record ID	3ZJ3
		Record imported (Y/N)	Y	Source Calibration			:LAPTOP COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/11/2009 03:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 14467	Download date/time: 01/16/2009	11:18	Record ID 3ZT8
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012321

Accuracy check date/time: 01/16/2009 11:18

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3736
Check sum 43790	Download date/time: 01/16/2009	11:18	Record ID 3ZS7
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/16/2009 11:21

First calibration check: .138

Interferent detected: No

Second calibration check: .140

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 024

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 44591	Download date/time: 01/16/2009	11:20	Record ID 3ZU1
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/17/2009 06:23

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 43723	Download date/time: 01/23/2009	10:03	Record ID 4032
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/17/2009 18:46

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16797	Download date/time: 01/23/2009	10:03	Record ID 4041
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012321

Accuracy check date/time: 01/20/2009 00:16

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 67870	Download date/time: 01/23/2009	10:03	Record ID 4052
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/22/2009 22:39

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3757
Check sum	37739	Download date/time:	01/23/2009 10:03	Record ID	4014		
Record imported (Y/N)			Y	Source Diagnostic	:LAPTOP COBRA		

Instrument serial No: 68-012321

Accuracy check date/time: 01/23/2009 09:53

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3769
Check sum	53173	Download date/time:	01/23/2009 10:03	Record ID	4021		
Record imported (Y/N)			Y	Source Diagnostic	:LAPTOP COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/23/2009 09:57

First calibration check: .137

Interferent detected: No

Second calibration check: .138

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 033

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 38765	Download date/time: 01/23/2009	10:03	Record ID 4003
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/23/2009 22:37

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Radio frequency interference detected

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 46585	Download date/time: 01/30/2009	11:45	Record ID 40J0
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/25/2009 23:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3706
Check sum	44838	Download date/time:	01/30/2009 11:45	Record ID	40G9		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012321 Accuracy check date/time: 01/26/2009 05:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3787
Check sum	91230	Download date/time:	01/30/2009 11:45	Record ID	40H0		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/27/2009 02:07

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 19903	Download date/time: 01/30/2009	11:45	Record ID 40K5
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/27/2009 02:16

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Radio frequency interference detected

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 39738	Download date/time: 01/30/2009	11:45	Record ID 40L2
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/29/2009 03:32

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 48821	Download date/time: 01/30/2009	11:45	Record ID 40M7
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012321

Accuracy check date/time: 01/29/2009 12:52

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Radio frequency interference detected

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 36140	Download date/time: 01/30/2009	11:45	Record ID 40N2
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/29/2009 12:55

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Radio frequency interference detected

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 13600	Download date/time: 01/30/2009	11:45	Record ID 4007
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/30/2009 11:34

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3789
Check sum 64091	Download date/time: 01/30/2009	11:45	Record ID 4015				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 01/30/2009 11:39

First calibration check: .140

Interferent detected: No

Second calibration check: .142

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C:
Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 036

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 98530	Download date/time: 01/30/2009	11:45	Record ID 40E9
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012321 Accuracy check date/time: 01/30/2009 11:41

First calibration check: .142

Interferent detected: No

Second calibration check: .143

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51328	Download date/time: 01/30/2009	11:45	Record ID 40F4
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321

Accuracy check date/time: 01/31/2009 00:01

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3769
Check sum	13663	Download date/time:	02/06/2009 09:44			Record ID	4155
		Record imported (Y/N)	Y	Source Diagnostic			:LAPTOP COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.