

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334

Accuracy check date/time: 01/02/2009 09:11

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3325
Check sum	62674	Download date/time:	01/02/2009	09:12		Record ID	3ZG0
	Record imported (Y/N)	Y	Source Diagnostic	:LAPTOP COBRA			

---

Instrument serial No: 68-012334

Accuracy check date/time: 01/02/2009 09:13

First calibration check: .115

Interferent detected: No

Second calibration check: .117

Interferent detected: No

Concentration (%W/V) .124

Solution lot#: IRS020708

Soln temp C: 34.03

Simulator Ser No. DR2492

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	37349	Download date/time:	01/02/2009	09:14		Record ID	3ZH7
	Record imported (Y/N)	Y	Source Calibration	:LAPTOP COBRA			

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334

Accuracy check date/time: 01/02/2009 09:13

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3314
Check sum	13514	Download date/time:	01/02/2009 09:14	Record ID	3Z14		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012334

Accuracy check date/time: 01/09/2009 10:52

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3283
Check sum	38295	Download date/time:	01/09/2009 10:53	Record ID	3ZQ1		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 01/09/2009 10:57

First calibration check: .132

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 33.99  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 29537	Download date/time: 01/09/2009	10:59	Record ID 3ZR4
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 01/16/2009 11:36

First calibration check: .133

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 58021	Download date/time: 01/16/2009	11:38	Record ID 3ZY1
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave  
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 01/16/2009 11:38

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass  
Prom Pass Real Time clock Pass Motor Pass Printer Pass  
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 3284  
Check sum 32315 Download date/time: 01/16/2009 11:38 Record ID 3ZZ4  
Record imported (Y/N) Y Source Diagnostic :LAPTOP COBRA

---

Instrument serial No: 68-012334 Accuracy check date/time: 01/23/2009 10:18

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass  
Prom Pass Real Time clock Pass Motor Pass Printer Pass  
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 3246  
Check sum 58179 Download date/time: 01/23/2009 10:22 Record ID 40D4  
Record imported (Y/N) Y Source Diagnostic :LAPTOP COBRA

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 01/23/2009 10:22

First calibration check: .129

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 10816	Download date/time: 01/23/2009	10:22	Record ID 40C7
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 01/30/2009 12:28

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3240
Check sum 27936	Download date/time: 01/30/2009	12:30	Record ID 4117
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave  
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 01/30/2009 12:30

First calibration check: .131

Interferent detected: No

Second calibration check: .135

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.01  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22849	Download date/time: 01/30/2009	12:30	Record ID 4108
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.