

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339 Accuracy check date/time: 01/23/2009 12:54

First calibration check: .138

Interferent detected: No

Second calibration check: .138

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.03
Simulator Ser No. DR3707

LPACED INTO SERVICE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 47485 | Download date/time: 01/30/2009 | 12:12 | Record ID 40P9 |
| Record imported (Y/N) Y | Source Calibration | :LAPTOP COBRA | |

Instrument serial No: 68-012339 Accuracy check date/time: 01/23/2009 14:03

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | | | | | |
|-------------------------|--------------------------------|------------------|----------------|-------|------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3366 |
| Check sum 60002 | Download date/time: 01/30/2009 | 12:12 | Record ID 40T1 | | | | |
| Record imported (Y/N) Y | Source Diagnostic | :LAPTOP COBRA | | | | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339 Accuracy check date/time: 01/24/2009 21:40

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 62560 | Download date/time: 01/30/2009 | 12:12 | Record ID 40U6 |
| Record imported (Y/N) Y | Source Exception | :LAPTOP COBRA | |

Instrument serial No: 68-012339 Accuracy check date/time: 01/25/2009 03:29

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 30034 | Download date/time: 01/30/2009 | 12:12 | Record ID 40V1 |
| Record imported (Y/N) Y | Source Exception | :LAPTOP COBRA | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339 Accuracy check date/time: 01/27/2009 02:03

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 37166 | Download date/time: 01/30/2009 | 12:12 | Record ID 40W6 |
| Record imported (Y/N) Y | Source Exception | :LAPTOP COBRA | |

Instrument serial No: 68-012339 Accuracy check date/time: 01/27/2009 02:04

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 73060 | Download date/time: 01/30/2009 | 12:12 | Record ID 40X7 |
| Record imported (Y/N) Y | Source Exception | :LAPTOP COBRA | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339

Accuracy check date/time: 01/27/2009 02:13

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 62285 | Download date/time: 01/30/2009 | 12:12 | Record ID 40Y2 |
| Record imported (Y/N) Y | Source Exception | :LAPTOP COBRA | |

Instrument serial No: 68-012339

Accuracy check date/time: 01/30/2009 12:04

First calibration check: .119

Interferent detected: No

Second calibration check: .119

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 050

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 43800 | Download date/time: 01/30/2009 | 12:12 | Record ID 40Q4 |
| Record imported (Y/N) Y | Source Calibration | :LAPTOP COBRA | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339 Accuracy check date/time: 01/30/2009 12:07

First calibration check: .119

Interferent detected: No

Second calibration check: .120

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.00
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 68310 | Download date/time: 01/30/2009 | 12:12 | Record ID 40R9 |
| Record imported (Y/N) Y | Source Calibration | :LAPTOP COBRA | |

Instrument serial No: 68-012339 Accuracy check date/time: 01/30/2009 12:07

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

TAKEN OUT OF SERVICE - READS LOW

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 13718 | Download date/time: 01/30/2009 | 12:12 | Record ID 40Z7 |
| Record imported (Y/N) Y | Source Memo | :LAPTOP COBRA | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012339

Accuracy check date/time: 01/30/2009 12:09

First calibration check: .121

Interferent detected: No

Second calibration check: .122

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.05

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 14530 | Download date/time: 01/30/2009 | 12:12 | Record ID 40S4 |
| Record imported (Y/N) Y | Source Calibration | :LAPTOP COBRA | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.