

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/05/2009 01:21

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3485
Check sum	13821	Download date/time:	02/06/2009 09:45	Record ID	41E8		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012335 Accuracy check date/time: 02/06/2009 09:39

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3445
Check sum	23576	Download date/time:	02/06/2009 09:45	Record ID	41F7		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/06/2009 09:52

First calibration check: .137

Interferent detected: No

Second calibration check: .139

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122609

Soln temp C: 33.99
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 029

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 42621	Download date/time: 02/06/2009	09:53	Record ID 41J5
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012335 Accuracy check date/time: 02/08/2009 02:13

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3443
Check sum 16467	Download date/time: 02/12/2009	12:13	Record ID 41U1
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/12/2009 12:05

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass
Prom Pass Real Time clock Pass Motor Pass Printer Pass
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 3456
Check sum 58375 Download date/time: 02/12/2009 12:13 Record ID 41V0
Record imported (Y/N) Y Source Diagnostic :LAPTOP COBRA

Instrument serial No: 68-012335 Accuracy check date/time: 02/12/2009 12:12

First calibration check: .139 Interferent detected: No
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122609 Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 019

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom Serial No. Temp Internal standards
Prom Real Time clock Motor Printer
Ram Auto calibration Range Digital voltmeter
Check sum 26350 Download date/time: 02/12/2009 12:13 Record ID 41T6
Record imported (Y/N) Y Source Calibration :LAPTOP COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.