

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 02/17/2009 12:15

First calibration check: .135

Interferent detected: No

Second calibration check: .136

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.03
Simulator Ser No. DR3707

PLACED INTO SERVICE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 59900	Download date/time: 02/19/2009	13:57	Record ID 4214
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012340 Accuracy check date/time: 02/17/2009 14:12

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3522
Check sum 56541	Download date/time: 02/19/2009	13:57	Record ID 4236
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/18/2009 20:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64414	Download date/time: 02/19/2009	13:57	Record ID 4249
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

Instrument serial No: 68-012340

Accuracy check date/time: 02/19/2009 13:54

First calibration check: .129

Interferent detected: No

Second calibration check: .133

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.07

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 007

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 25492	Download date/time: 02/19/2009	13:57	Record ID 4221
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 02/27/2009 12:25

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3274
Check sum	78370	Download date/time:	02/27/2009	12:31		Record ID	42V5
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012340 Accuracy check date/time: 02/27/2009 12:31

First calibration check: .128

Interferent detected: No

Second calibration check: .130

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.05

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 053

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards			
Prom	Real Time clock	Motor	Printer			
Ram	Auto calibration	Range	Digital voltmeter			
Check sum	26750	Download date/time:	02/27/2009	12:33	Record ID	42W6
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.