

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 03/05/2009 17:20

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3051
Check sum	15224	Download date/time:	03/05/2009	17:20		Record ID	42X3
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012334 Accuracy check date/time: 03/05/2009 17:23

First calibration check: .131

Interferent detected: No

Second calibration check: .132

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.04

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	28088	Download date/time:	03/05/2009	17:22		Record ID	42Y4
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 03/08/2009 05:43

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

CHANGED TIME FROM 0543 TO 0643

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 54103	Download date/time: 03/13/2009	12:53	Record ID 4355
Record imported (Y/N) Y	Source Memo	:LAPTOP COBRA	

Instrument serial No: 68-012334 Accuracy check date/time: 03/13/2009 12:51

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3151
Check sum 42713	Download date/time: 03/13/2009	12:53	Record ID 4340
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 03/13/2009 12:53

First calibration check: .130

Interferent detected: No

Second calibration check: .132

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 003

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 45580	Download date/time: 03/13/2009	12:54	Record ID 4366
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012334 Accuracy check date/time: 03/19/2009 11:27

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3107
Check sum 32498	Download date/time: 03/19/2009	11:29	Record ID 43Q5
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 03/19/2009 11:30

First calibration check: .130

Interferent detected: No

Second calibration check: .132

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.04
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 62206	Download date/time: 03/19/2009	11:30	Record ID 43R2
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

Instrument serial No: 68-012334 Accuracy check date/time: 03/26/2009 10:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3083
Check sum 55431	Download date/time: 03/26/2009	11:01	Record ID 4456				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012334 Accuracy check date/time: 03/26/2009 11:01

First calibration check: .134

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.04

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12950	Download date/time: 03/26/2009	11:01	Record ID 4445
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.