

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/02/2009 15:33

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3341
Check sum	57607	Download date/time:	04/02/2009	15:40		Record ID	44T9
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

---

Instrument serial No: 68-012340

Accuracy check date/time: 04/02/2009 15:38

First calibration check: .128

Interferent detected: No

Second calibration check: .129

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.01

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 044

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	51324	Download date/time:	04/02/2009	15:40		Record ID	44S8
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/04/2009 03:38

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 43744	Download date/time: 04/09/2009	10:29	Record ID 4575
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

---

Instrument serial No: 68-012340

Accuracy check date/time: 04/09/2009 10:23

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3338
Check sum 24370	Download date/time: 04/09/2009	10:29	Record ID 4566				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave  
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/09/2009 10:28

First calibration check: .130

Interferent detected: No

Second calibration check: .132

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.07  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 049

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 27410	Download date/time: 04/09/2009	10:30	Record ID 4584
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012340 Accuracy check date/time: 04/13/2009 03:29

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31610	Download date/time: 04/17/2009	10:48	Record ID 45J5
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/17/2009 10:37

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3369
Check sum	31829	Download date/time:	04/17/2009	10:48		Record ID	4514
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

Instrument serial No: 68-012340 Accuracy check date/time: 04/17/2009 10:44

First calibration check: .128

Interferent detected: No

Second calibration check: .133

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 054

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards		
Prom	Real Time clock	Motor	Printer		
Ram	Auto calibration	Range	Digital voltmeter		
Check sum	52220	Download date/time:	04/17/2009 10:48	Record ID	45H9
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/17/2009 17:38

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3359
Check sum	47058	Download date/time:	04/23/2009 14:06	Record ID	45Z8		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012340

Accuracy check date/time: 04/21/2009 15:20

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3371
Check sum	19100	Download date/time:	04/23/2009 14:06	Record ID	4605		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/23/2009 14:00

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3366
Check sum	25302	Download date/time:	04/23/2009	14:06		Record ID	4610
Record imported (Y/N)		Y	Source Diagnostic		:LAPTOP COBRA		

---

Instrument serial No: 68-012340

Accuracy check date/time: 04/23/2009 14:05

First calibration check: .129

Interferent detected: No

Second calibration check: .130

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.04

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 012

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	26449	Download date/time:	04/23/2009	14:06		Record ID	45Y7
Record imported (Y/N)		Y	Source Calibration		:LAPTOP COBRA		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/30/2009 02:46

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 58128	Download date/time: 04/30/2009	12:02	Record ID 46D0
Record imported (Y/N) Y	Source Exception	:LAPTOP COBRA	

---

Instrument serial No: 68-012340

Accuracy check date/time: 04/30/2009 11:53

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3371
Check sum 33378	Download date/time: 04/30/2009	12:02	Record ID 46C7
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340

Accuracy check date/time: 04/30/2009 11:59

First calibration check: .132

Interferent detected: No

Second calibration check: .132

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 034

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 50759	Download date/time: 04/30/2009	12:02	Record ID 46B8
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.