

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/02/2009 02:22

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |                   |       |               |                    |      |
|-----------------------|-------|---------------------|-------------------|-------|---------------|--------------------|------|
| Eeprom                | Pass  | Serial No.          | Pass              | Temp  | Pass          | Internal standards | Pass |
| Prom                  | Pass  | Real Time clock     | Pass              | Motor | Pass          | Printer            | Pass |
| Ram                   | Pass  | Auto calibration    | Pass              | Range | Pass          | Digital voltmeter  | 4056 |
| Check sum             | 27992 | Download date/time: | 05/08/2009        | 10:44 |               | Record ID          | 46N2 |
| Record imported (Y/N) |       | Y                   | Source Diagnostic |       | :LAPTOP COBRA |                    |      |

Instrument serial No: 68-012321

Accuracy check date/time: 05/03/2009 19:39

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |                  |       |               |                    |      |
|-----------------------|-------|---------------------|------------------|-------|---------------|--------------------|------|
| Eeprom                |       | Serial No.          |                  | Temp  |               | Internal standards |      |
| Prom                  |       | Real Time clock     |                  | Motor |               | Printer            |      |
| Ram                   |       | Auto calibration    |                  | Range |               | Digital voltmeter  |      |
| Check sum             | 60439 | Download date/time: | 05/08/2009       | 10:44 |               | Record ID          | 46P5 |
| Record imported (Y/N) |       | Y                   | Source Exception |       | :LAPTOP COBRA |                    |      |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/07/2009 04:27

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 60236         | Download date/time: 05/08/2009 | 10:44         | Record ID 46Q6     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

Instrument serial No: 68-012321

Accuracy check date/time: 05/07/2009 04:29

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 57381         | Download date/time: 05/08/2009 | 10:44         | Record ID 46R3     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/08/2009 10:39

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|           |       |                       |            |                   |      |                    |               |
|-----------|-------|-----------------------|------------|-------------------|------|--------------------|---------------|
| Eeprom    | Pass  | Serial No.            | Pass       | Temp              | Pass | Internal standards | Pass          |
| Prom      | Pass  | Real Time clock       | Pass       | Motor             | Pass | Printer            | Pass          |
| Ram       | Pass  | Auto calibration      | Pass       | Range             | Pass | Digital voltmeter  | 4064          |
| Check sum | 12539 | Download date/time:   | 05/08/2009 | 10:44             |      | Record ID          | 46O1          |
|           |       | Record imported (Y/N) | Y          | Source Diagnostic |      |                    | :LAPTOP COBRA |

---

Instrument serial No: 68-012321

Accuracy check date/time: 05/08/2009 10:48

First calibration check: .143

Interferent detected: No

Second calibration check: .144

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.01

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 043

Operator G GOSSAGE 7885

---

**Diagnostics check**

|           |       |                       |            |                    |  |                    |               |
|-----------|-------|-----------------------|------------|--------------------|--|--------------------|---------------|
| Eeprom    |       | Serial No.            |            | Temp               |  | Internal standards |               |
| Prom      |       | Real Time clock       |            | Motor              |  | Printer            |               |
| Ram       |       | Auto calibration      |            | Range              |  | Digital voltmeter  |               |
| Check sum | 39423 | Download date/time:   | 05/08/2009 | 10:48              |  | Record ID          | 46T5          |
|           |       | Record imported (Y/N) | Y          | Source Calibration |  |                    | :LAPTOP COBRA |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/09/2009 02:23

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 47567         | Download date/time: 05/14/2009 | 16:19         | Record ID 4746     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

Instrument serial No: 68-012321

Accuracy check date/time: 05/09/2009 02:51

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 58244         | Download date/time: 05/14/2009 | 16:19         | Record ID 4751     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/12/2009 04:22

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 18229         | Download date/time: 05/14/2009 | 16:19         | Record ID 4760     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

Instrument serial No: 68-012321

Accuracy check date/time: 05/13/2009 02:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 41032         | Download date/time: 05/14/2009 | 16:19         | Record ID 4775     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321

Accuracy check date/time: 05/13/2009 07:10

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |               |                    |
|-------------------------|--------------------------------|---------------|--------------------|
| Eeprom                  | Serial No.                     | Temp          | Internal standards |
| Prom                    | Real Time clock                | Motor         | Printer            |
| Ram                     | Auto calibration               | Range         | Digital voltmeter  |
| Check sum 31285         | Download date/time: 05/14/2009 | 16:19         | Record ID 4782     |
| Record imported (Y/N) Y | Source Exception               | :LAPTOP COBRA |                    |

---

Instrument serial No: 68-012321

Accuracy check date/time: 05/14/2009 16:17

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

|                         |                                |                  |                |       |      |                       |
|-------------------------|--------------------------------|------------------|----------------|-------|------|-----------------------|
| Eeprom                  | Pass                           | Serial No.       | Pass           | Temp  | Pass | Internal standards    |
| Prom                    | Pass                           | Real Time clock  |                | Motor | Pass | Printer               |
| Ram                     | Pass                           | Auto calibration | Pass           | Range | Fail | Digital voltmeter ??? |
| Check sum 17433         | Download date/time: 05/14/2009 | 16:19            | Record ID 4711 |       |      |                       |
| Record imported (Y/N) Y | Source Diagnostic              | :LAPTOP COBRA    |                |       |      |                       |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321 Accuracy check date/time: 05/14/2009 16:18

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |            |                   |               |                       |
|-----------------------|-------|---------------------|------------|-------------------|---------------|-----------------------|
| Eeprom                | Pass  | Serial No.          | Pass       | Temp              | Pass          | Internal standards    |
| Prom                  | Pass  | Real Time clock     |            | Motor             | Pass          | Printer               |
| Ram                   | Pass  | Auto calibration    | Pass       | Range             | Fail          | Digital voltmeter ??? |
| Check sum             | 20380 | Download date/time: | 05/14/2009 | 16:19             |               | Record ID 4720        |
| Record imported (Y/N) |       |                     | Y          | Source Diagnostic | :LAPTOP COBRA |                       |

Instrument serial No: 68-012321 Accuracy check date/time: 05/14/2009 16:20

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |            |                   |               |                       |
|-----------------------|-------|---------------------|------------|-------------------|---------------|-----------------------|
| Eeprom                | Pass  | Serial No.          | Pass       | Temp              | Pass          | Internal standards    |
| Prom                  | Pass  | Real Time clock     |            | Motor             | Pass          | Printer               |
| Ram                   | Pass  | Auto calibration    | Pass       | Range             | Fail          | Digital voltmeter ??? |
| Check sum             | 56780 | Download date/time: | 05/14/2009 | 16:19             |               | Record ID 4735        |
| Record imported (Y/N) |       |                     | Y          | Source Diagnostic | :LAPTOP COBRA |                       |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.