

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave  
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 06/08/2009 06:58

First calibration check: .129

Interferent detected: No

Second calibration check: .130

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.02  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 46740	Download date/time: 06/08/2009	06:57	Record ID 48B9
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 06/08/2009 07:00

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 2869
Check sum 59437	Download date/time: 06/12/2009	08:54	Record ID 48O6
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 06/12/2009 08:54

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass  
Prom Pass Real Time clock Pass Motor Pass Printer Pass  
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 2885  
Check sum 51980 Download date/time: 06/12/2009 08:54 Record ID 48P8  
Record imported (Y/N) Y Source Diagnostic :LAPTOP COBRA

---

Instrument serial No: 68-012334 Accuracy check date/time: 06/12/2009 08:59

First calibration check: .128 Interferent detected: No  
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608 Soln temp C: 34.02  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom Serial No. Temp Internal standards  
Prom Real Time clock Motor Printer  
Ram Auto calibration Range Digital voltmeter  
Check sum 17139 Download date/time: 06/12/2009 09:01 Record ID 48Q1  
Record imported (Y/N) Y Source Calibration :LAPTOP COBRA

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334

Accuracy check date/time: 06/18/2009 20:38

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2801
Check sum	48319	Download date/time:	06/19/2009 10:52	Record ID	48S1		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

Instrument serial No: 68-012334

Accuracy check date/time: 06/19/2009 10:49

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2873
Check sum	50130	Download date/time:	06/19/2009 10:52	Record ID	48T6		
Record imported (Y/N)		Y	Source Diagnostic	:LAPTOP COBRA			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave  
Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334 Accuracy check date/time: 06/19/2009 10:53

First calibration check: .133

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138 Solution lot#: IRS122608

Soln temp C: 34.05  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 18876	Download date/time: 06/19/2009	10:55	Record ID 48U9
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

Instrument serial No: 68-012334 Accuracy check date/time: 06/25/2009 13:40

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2853
Check sum 64113	Download date/time: 06/25/2009	13:43	Record ID 48Z0				
Record imported (Y/N) Y	Source Diagnostic	:LAPTOP COBRA					

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Crime Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012334

Accuracy check date/time: 06/25/2009 13:44

First calibration check: .130

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .138

Solution lot#: IRS122608

Soln temp C: 34.02

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49270	Download date/time: 06/25/2009	13:45	Record ID 4901
Record imported (Y/N) Y	Source Calibration	:LAPTOP COBRA	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.